

12025 NE Marx St. Portland, OR 97220 503-253-3511 / www.greenleaflabs.com License#: 10029074C70

TRF22010-0102

Sample ID: G2L0156-04

Matrix: Hemp Products

Test ID: 5026216 Source ID:

Date Sampled: 12/10/22

Date Accepted: 12/10/22

Harvest/Prod. Date: 12.05.2022

Periodic Caramels

shane@periodiccaramels.com

Total THC: 0.1791 %

Total CBD: 4.192 %

Total CBG: 1.082 %





Eric Wendt Chief Science Officer - 12/13/2022



12025 NE Marx St. Portland, OR 97220 503-253-3511 / www.greenleaflabs.com

License#: 10029074C70

TRF22010-0102

Sample ID: G2L0156-04

Matrix: Hemp Products

Test ID: 5026216 Source ID:

Date Sampled: 12/10/22

Date Accepted: 12/10/22

Harvest/Prod. Date: 12.05.2022

Periodic Caramels

shane@periodiccaramels.com

			Potency Analysis	
Date/Time Extra	cted: 12/12/	22 13:29	Analysis Method/SOP: 215 Batch Identification: 2251	00
Cannabinoids	LOQ (%)	mg/g	Cannabinoids Profile	
Total THC	0.0009	1.791		
Total CBD	0.0090	41.92	11	
Total CBG	0.0248	10.82	0.1	
THCA	0.0009	< LOQ		
delta 9-THC	0.0009	1.791		
delta 8-THC	0.0344	< LOQ		
THCV	0.0268	< LOQ		
THCVA	0.0401	< LOQ		
CBD	0.0090	41.92		_
CBDA	0.0090	< LOQ	delta 9-TH ■ CBD	IC
CBDV	0.0276	< LOQ	■ CBG CBC	
CBDVA	0.0379	< LOQ	Total:	
CBN	0.0248	< LOQ		
CBG	0.0288	10.82	* / / / / /	
CBGA	0.0382	< LOQ	4.2	
СВС	0.0362	0.523		

Total THC = delta 9-THC + (THCA * 0.877)

Total CBD = CBD + (CBDA * 0.877)

Total CBG = CBG + (CBGA * 0.878)

LOQ=Limit of Quantification, the lowest measurable concentration of an analyte.





Eric Wendt
Chief Science Officer -

Chief Science Officer - 12/13/2022



12025 NE Marx St. Portland, OR 97220 503-253-3511 / www.greenleaflabs.com License#: 10029074C70

Quality Control Potency

Batch: 2251005 - 215-Products

Blank(2251005-BLK1)							
Analyte	Result	LOQ	Units	%Recovery Limits	Extracted	Analyzed	Notes
THCA	< LOQ	0.0080	%		12/12/22 13:29	12/12/22 22:40	
delta 9-THC	< LOQ	0.0080	%		12/12/22 13:29	12/12/22 22:40	
delta 8-THC	< LOQ	0.3082	%		12/12/22 13:29	12/12/22 22:40	
THCV	< LOQ	0.2405	%		12/12/22 13:29	12/12/22 22:40	
THCVA	< LOQ	0.3599	%		12/12/22 13:29	12/12/22 22:40	
CBD	< LOQ	0.0080	%		12/12/22 13:29	12/12/22 22:40	
CBDA	< LOQ	0.0080	%		12/12/22 13:29	12/12/22 22:40	
CBDV	< LOQ	0.2474	%		12/12/22 13:29	12/12/22 22:40	
CBDVA	< LOQ	0.3398	%		12/12/22 13:29	12/12/22 22:40	
CBN	< LOQ	0.2222	%		12/12/22 13:29	12/12/22 22:40	
CBG	< LOQ	0.2585	%		12/12/22 13:29	12/12/22 22:40	
CBGA	< LOQ	0.3426	%		12/12/22 13:29	12/12/22 22:40	
CBC	< LOQ	0.3242	%		12/12/22 13:29	12/12/22 22:40	

Reference(2251005-SRM1)							
Analyte	% Recovery	LOQ	Units	%Recovery Limits	Extracted	Analyzed	Notes
THCA	90.1	0.0099	%	90-110	12/12/22 13:29	12/13/22 11:42	
delta 9-THC	99.8	0.0099	%	90-110	12/12/22 13:29	12/13/22 11:42	
delta 8-THC	97.0	0.3797	%	90-110	12/12/22 13:29	12/13/22 11:42	
CBD	99.5	0.0099	%	90-110	12/12/22 13:29	12/13/22 11:42	
CBDA	98.1	0.0099	%	90-110	12/12/22 13:29	12/13/22 11:42	







12025 NE Marx St. Portland, OR 97220 503-253-3511 / www.greenleaflabs.com

License#: 10029074C70

Notes and Definitions

Regulatory Compliance samples were collected onsite at facility according to ORELAP-SOP-001 and ORELAP-SOP-002 and following Sampling Plan FN117. Quality Control samples were tested as received. Results do not include uncertainty of measurements. Available upon request.

ATM	Non-cannabis matrix related interference or suppression of Internal standard
BLI	Baseline Interference - Cannabinoid peak interference in chromatographic baseline affecting QC recovery.
BLK	Analyte detected in method blank, but not associated samples.
BSH	Blank Spike High - Blank Spike recovery above method limit. no detections in samples.
BSL	Blank Spike Low - Blank Spike recovery below lower method limit, analyte chromatography reviewed
С	manually for all samples.
CBD	Interference due to co-elution
CV1	CBD matrix interference on GC Pest chromatography
CV2	CCV was above acceptance criteria, Non-detect samples are considered acceptable.
INF	CCV was below acceptance criteria, sample still exceeds regulatory limit.
ISH	One or more QC falls outside acceptance criteria. Data entered into LIMS for informational purposes only.
ISL	Internal Standard concentration is above acceptance criteria.
MSH	Internal Standard concentration is below acceptance criteria.
MSI	Matrix Spike High - Matrix Spike recovery above method limits.
MSL	Matrix Spike Interference - Matrix spike source sample contains analyte hit above calibration affecting
TPP	recovery accuracy in Matrix Spike.
U	Matrix Spike Low - Matrix Spike recovery below lower method limit, analyte chromatography reviewed

manually for all samples.

Internal Standard concentration outside control limit due to matrix interference



